

CERTIFICATE OF ANALYSIS

Product name: **BIOWAX PLUS 56/58**

Batch: BWPLUS-18/20

Expiry date: 31.05.25.

Product description

- Catalogue number: BWPLUS-1, BWPLUS-2, BWPLUS-10
- Product name: BIOWAX PLUS 56/58

Application

- Histology

Chemical and physical properties

- Aggregate state: solid
- Appearance: granules
- Solubility (20 °C): insoluble in water

Test	Method	Result	Specification
Congealing point	ASTM D1321	55 °C	54 – 57 °C
Visual (Color Saybolt)	ASTM D156	30	+25 min
Flocculation		Pass	Pass

Storage

- Store at +15 °C to +25 °C.

QA result

- PASSES TEST

We declare that the above mentioned serial number was produced in accordance with the technical specifications and has been tested in the diagnostic laboratory meeting all standard requirements.

BioGnost Ltd.